Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Chromium Standard, 1 mg/L Aqueous**
Catalog Number:	C1104
Lot Number:	C11040692501
Manufacture Date:	03/10/25
Expiration Date:	03/10/26

Property	Specification	Result
Concentration, Chromium	1.00 ± 0.02 mg/L	Conforms

Prepared from 1000 ppm AA chromium standard (certified traceable to NIST), and deionized water. Dilutions were made with Class A volumetric glassware.

Certificate approved on 03/10/25 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.