Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Nickel Standard, 1 mg/L in 5% v/v Nitric Acid**
Catalog Number:	N1039
Lot Number:	N10390222601
Manufacture Date:	01/22/26
Expiration Date:	01/22/27

Property	Specification	Result
Concentration, Nickel	1.00 ± 0.02 mg/L	Conforms

Prepared from 1000 ppm AA nickel standard (certified traceable to NIST), ACS Reagent Grade nitric acid, and deionized water. Dilutions were made with Class A volumetric glassware.

Certificate approved on 01/22/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.