Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951


CERTIFICATE OF ANALYSIS

Product:	**EDTA, 0.010 Molar**
Catalog Number:	E1005
Lot Number:	E10051112601
Manufacture Date:	04/21/26
Expiration Date:	04/21/28

Property	Specification	Result
Assay, EDTA	0.0100 ± 0.0002 M	0.0100 M

Prepared with ACS Reagent Grade ethylenediaminetetraacetic acid, disodium salt, dihydrate, and high purity water. Standardized against calcium carbonate standard certified traceable to NIST. Indicator: Eriochrome Black T. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 04/22/26 by:



Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.