Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS


Product:	**Hydrochloric Acid, 0.100 Normal**
Catalog Number:	H1008
Lot Number:	H10081532601
Manufacture Date:	06/02/26
Expiration Date:	06/02/28

Property	Specification	Result
Appearance	Clear, colorless liquid	Pass
Assay, HCl	0.1000 ± 0.0008 N	0.1001 N

Conforms to RMR Specification QA-PD-H1008 Revision A.

Prepared with ACS Reagent Grade hydrochloric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 06/03/26 by:



Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.