Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Sodium Thiosulfate, 0.02 Normal**
Catalog Number:	S1038
Lot Number:	S10381612601
Manufacture Date:	06/10/26
Expiration Date:	06/10/28

Property	Specification	Result
Assay, $Na_2S_2O_3$	0.0200 ± 0.0002 N	Conforms*

Prepared with ACS Reagent Grade sodium thiosulfate, pentahydrate, and high purity water. Contains < 0.1% sodium hydroxide as a stabilizer. *Note: This solution is subject to chemical degradation and should be restandardized prior to use for accurate work.

Certificate approved on 06/10/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com