Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Ceric Ammonium Sulfate, 0.100 Normal in 2.0 Normal Sulfuric Acid**
Catalog Number:	C1015
Lot Number:	C10151822601
Manufacture Date:	07/01/26
Expiration Date:	07/01/28

Property	Specification	Result
Assay, $Ce(NH_4)_4(SO_4)_4$	0.100 ± 0.003 N	0.100 N

Conforms to RMR Specification QA-PD-C1015 Revision A.

Prepared with ACS Reagent Grade ceric ammonium sulfate, dihydrate, ACS Reagent Grade sulfuric acid, and high purity water. Standardized against ferrous ammonium sulfate solution certified traceable to NIST. Indicator: ferroin. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 07/06/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.