Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Silver Nitrate, 0.010 Normal**
Catalog Number:	S1070
Lot Number:	S10701692601
Manufacture Date:	06/18/26
Expiration Date:	06/18/28

Property	Specification	Result
Assay, $AgNO_3$	0.0100 ± 0.0001 N	0.0100 N

Prepared with ACS Reagent Grade silver nitrate and high purity water. Standardized against potassium thiocyanate solution certified traceable to NIST. Indicator: ferric ammonium sulfate. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 09/28/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com