Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Hydrochloric Acid, 0.250 Normal**
Catalog Number:	H1002
Lot Number:	H10022242601
Manufacture Date:	08/12/26
Expiration Date:	08/12/28

Property	Specification	Result
Assay, HCl	0.250 ± 0.005 N	0.250 N

Prepared with ACS Reagent Grade hydrochloric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final QC release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 08/13/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com