Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**EDTA, 0.100 Molar**
Catalog Number:	E1008
Lot Number:	E10082302601
Manufacture Date:	08/18/26
Expiration Date:	08/18/28

Property	Specification	Result
Appearance	Clear, colorless liquid	Pass
Assay, EDTA	0.100 ± 0.003 M	0.100 M

Conforms to RMR Specification QA-PD-E1008 Revision B.

Prepared with ACS Reagent Grade ethylenediaminetetraacetic acid, disodium salt, dihydrate, and high purity water. Standardized against calcium carbonate standard certified traceable to NIST. Indicator: Eriochrome Black T. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 08/18/26 by:



Matt Jones
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.