Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Hydrochloric Acid, 6.00 Normal**
Catalog Number:	H1013
Lot Number:	H10132322601
Manufacture Date:	08/20/26
Expiration Date:	08/20/28

Property	Specification	Result
Appearance	Clear, colorless liquid	Pass
Assay, HCl	6.00 ± 0.05 N	6.00 N

Conforms to RMR Specification QA-PD-H1013 Revision A.

Prepared with ACS Reagent Grade hydrochloric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 08/24/26 by:



Matt Jones
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.