Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**EDTA, 0.0575 Molar**
Catalog Number:	E1007
Lot Number:	E10072392601
Manufacture Date:	08/27/26
Expiration Date:	08/27/28

Property	Specification	Result
Assay, EDTA	0.0575 ± 0.0008 M	0.0576 M

Prepared with ACS Reagent Grade ethylenediaminetetraacetic acid, disodium salt, dihydrate, and high purity water. Standardized against calcium carbonate standard certified traceable to NIST. Indicator: Eriochrome Black T. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 08/27/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com