Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Hydrochloric Acid**, **3.00 Normal**
Catalog Number:	H2008
Lot Number:	H20082522601
Manufacture Date:	09/09/26
Expiration Date:	09/09/28

Property	Specification	Result
Assay, HCl	3.00 ± 0.02 N	3.00 N

Prepared with ACS Reagent Grade hydrochloric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 09/10/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.