Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Iodine, 0.100 Normal**
Catalog Number:	I1010
Lot Number:	I10102532601
Manufacture Date:	09/10/26
Expiration Date:	09/10/28

Property	Specification	Result
Appearance	Yellowish brown liquid	Pass
Assay, I_2	0.1000 ± 0.0008 N*	0.0999 N*

Conforms to RMR Specification QA-PD-I1010 Revision B.

Prepared with ACS Reagent Grade potassium iodide, ACS Reagent Grade iodine, and high purity water. Standardized against sodium thiosulfate solution certified traceable to NIST. Indicator: starch solution. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value. *Note: This solution is subject to chemical degradation and should be restandardized prior to use for accurate work.

Certificate approved on 09/30/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com