Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Hydrochloric Acid, 5.00 Normal**
Catalog Number:	H1012
Lot Number:	H10122542601
Manufacture Date:	09/11/26
Expiration Date:	09/11/28

Property	Specification	Result
Assay, HCl	5.00 ± 0.05 N	5.01 N

Prepared with ACS Reagent Grade hydrochloric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 09/11/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.