Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Iodine, 0.020 Normal**
Catalog Number:	I1017
Lot Number:	I10172662601
Manufacture Date:	09/23/26
Expiration Date:	09/23/28

Property	Specification	Result
Concentration, Iodine (I_2)	0.0200 ± 0.0005 N*	0.0201 N*

Prepared with ACS Reagent Grade potassium iodide, ACS Reagent Grade iodine, and high purity water. Standardized against sodium thiosulfate solution certified traceable to NIST. Indicator: starch solution. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value. *Note: This solution is subject to chemical degradation and should be restandardized prior to use for accurate work.

Certificate approved on 09/23/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.