Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Silver Nitrate, 0.153 Normal**
Catalog Number:	S1006
Lot Number:	S10062682601
Manufacture Date:	09/25/26
Expiration Date:	09/25/28

Property	Specification	Result
Assay, $AgNO_3$	0.153 ± 0.001 N	Conforms

Prepared with ACS Reagent Grade silver nitrate and high purity water. Standardized against potassium thiocyanate solution certified traceable to NIST. Indicator: ferric ammonium sulfate. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 09/25/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.

4621 Technology Drive Golden, CO 80403
(303)762-0800 | f: (303)762-1240 | www.rmreagents.com | info@rmreagents.com