Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**Nitric Acid, 0.100 Normal**
Catalog Number:	N1004
Lot Number:	N10042572601
Manufacture Date:	09/14/26
Expiration Date:	09/14/28

Property	Specification	Result
Assay, HNO_3	0.100 ± 0.001 N	0.100 N

Prepared with ACS Reagent Grade nitric acid and high purity water. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 09/29/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.